

EC DECLARATION OF CONFORMITY

Moor Instruments Ltd
Millwey
Axminster
Devon
EX13 5HU
England

Declares that the medical device described hereafter

IRLD20 High Power Laser Source for DRT4 LDF Monitor,

is in conformity with the essential requirements and provisions of Council Directive 93/42/EEC and is in conformity with the national standard transposing harmonised standards.

EN 60601-1: 1990

EN 60825-1: 1994 incorporating amendment Nos 1, 2 and 3

is subject to the procedure set out in Annex 2 of Directive 93/42/EEC under the supervision of Notified Body Number 0120, SGS United Kingdom Limited, Unit 202B, Worle Parkway, Weston-Super-Mare, Somerset BS22 6WA.

Axminster, England 19/06/2000

Managing Director
on behalf of Moor Instruments Ltd