

# EC DECLARATION OF CONFORMITY

Moor Instruments Ltd  
Millwey  
Axminster  
Devon  
EX13 5HU  
UK

Declares that the medical device described hereafter **moorVMS-OXY Tissue Oxygen and Temperature Monitor** is in conformity with the essential requirements and provisions of Council Directive 93/42/EEC and is in conformity with the national standard transposing harmonised standards:

IEC 60601-1: 2005  
IEC 60601-1-2: 2007  
IEC 60825-1: 2001

Is subject to the procedure set out in Annex 2 of Directive 93/42/EEC under the supervision of Notified Body Number 0120, SGS United Kingdom Limited, Unit 202B, Worle Parkway, Weston-Super-Mare, Somerset BS22 6WA, UK

Axminster, England 06/10/2010



---

D. Boggett  
Managing Director  
On behalf of Moor Instruments Ltd

©2011 Moor Instruments Limited, Millwey, Axminster, Devon, EX13 5HU, UK.  
All rights reserved. Printed in England.

Manual Version: moorVMS-OXY User Manual (Version 2.0 English)  
Revision Date: 17/01/2011

This document may not, in whole or in part, be translated or reproduced by any means without prior consent from Moor Instruments Limited.

---